



SCS Directory

Accreditation number: SCS 0173

International standard: ISO/IEC 17025:2017

Swiss standard: SN EN ISO/IEC 17025:2018

EMC-Testcenter AG
Moosäckerstrasse 77
Postfach
8105 Regensdorf

Head: Marcel Fehr
Responsible for MS: Max Hunziker
Telephone: +41 44 302 45 00
E-Mail: calibration@emc-testcenter.com
Internet: www.emc-testcenter.com
Initial accreditation: 21.09.2023
Current accreditation: 21.09.2023 to 20.09.2028
Scope of accreditation see: www.sas.admin.ch
(Accredited bodies)

Scope of accreditation as of 21.09.2023

Calibration laboratory for electrical measurands

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Scattering parameters (complex): Reflection coefficient in coaxial line (S11)	0,01 ... 1 0,0001 ... 1	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz 9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	0,0246 0,0119 0,0171 0,0276 0,0353 0,0293 0,0263 0,0287 0,0401 0,0497	$Z_{ref} = 50 \Omega$, N-connector, real and imaginary quantity Including additional uncertainty for connector and cable movement according EURAMET Calibration Guide No. 12, Version 3.0



SCS Directory

Accreditation number: SCS 0173

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Scattering parameters (complex): Transmission coefficient in coaxial line (S21)	0.01 ... 1 (40 dB)	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	0,0247 (0,212 dB) 0,0121 (0,104 dB) 0,0193 (0,166 dB) 0,0276 (0,236 dB) 0,0355 (0,303 dB)	$Z_{ref} = 50 \Omega$, N-connector, real and imaginary quantity
	0.0001 ... 1 (80 dB)	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	0,0294 (0,252 dB) 0,0264 (0,227 dB) 0,0287 (0,246 dB) 0,0401 (0,342 dB) 0,0498 (0,422 dB)	Including additional uncertainty for connector and cable movement according EURAMET Calibration Guide No. 12, Version 3.0
Scattering parameters (derived quantity): RF Impedance	2 ... < 15 Ω	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	(-1,8E-2 · Z + 3,8E-1) · Z (-8,5E-3 · Z + 1,8E-1) · Z (-1,2E-2 · Z + 2,6E-1) · Z (-2,0E-2 · Z + 4,3E-1) · Z (-2,5E-2 · Z + 5,5E-1) · Z	Z = measured RF Impedance, $Z_{ref} = 50 \Omega$, N-connector
	15 ... < 100 Ω	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	6,1E-2 · Z 2,9E-2 · Z 4,2E-2 · Z 6,9E-2 · Z 8,8E-2 · Z	Including additional uncertainty for connector and cable movement according EURAMET Calibration Guide No. 12, Version 3.0
	100 Ω ... 1 k Ω	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	(3,2E-4 · Z + 2,5E-2) · Z (1,3E-4 · Z + 1,4E-2) · Z (2,0E-4 · Z + 1,9E-2) · Z (3,7E-4 · Z + 2,7E-2) · Z (5,2E-4 · Z + 3,0E-2) · Z	
Scattering parameters (derived quantity): RF Admittance	1 mS ... < 10 mS	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	(-2,0E+1 · Y + 2,4E-1) · Y (-9,6E+0 · Y + 1,2E-1) · Y (-1,4E+1 · Y + 1,7E-1) · Y (-2,2E+1 · Y + 2,7E-1) · Y (-2,9E+1 · Y + 3,5E-1) · Y	Y = measured RF Admittance, $Y_{ref} = 20$ mS, N-connector
	10 mS ... < 50 mS	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	6,3E-2 · Y 3,0E-2 · Y 4,3E-2 · Y 7,1E-2 · Y 9,2E-2 · Y	Including additional uncertainty for connector and cable movement according EURAMET Calibration Guide No. 12, Version 3.0
	50 mS ... 0.5 S	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	(9,8E-1 · Y + 1,7E-2) · Y (3,6E-1 · Y + 1,2E-2) · Y (5,8E-1 · Y + 1,5E-2) · Y (1,2E+0 · Y + 1,6E-2) · Y (1,9E+0 · Y + 1,1E-2) · Y	



SCS Directory

Accreditation number: SCS 0173

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Scattering parameters (derived quantity): Voltage Standing Wave Ratio (VSWR)	1,06 ... 10	9 kHz ... < 100 kHz 100 kHz ... < 1 GHz 1 GHz ... < 6 GHz 6 GHz ... < 12 GHz 12 GHz ... 18 GHz	(1,17E-2 · V + 3,80E-2) · V (0,53E-2 · V + 1,85E-2) · V (0,79E-2 · V + 2,65E-2) · V (1,34E-2 · V + 4,26E-2) · V (1,79E-2 · V + 5,44E-2) · V	V = measured VSWR $Z_{ref} = 50 \Omega$, N-connector Including additional uncertainty for connector and cable movement according EURAMET Calibration Guide No. 12, Version 3.0
CDN parameters:				
Impedance, Z_{CM} (common mode)	80 ... 220 Ω	150 kHz ... < 80 MHz 80 MHz ... < 200 MHz 200 MHz ... 230 MHz	5,0 Ω 7,0 Ω 25,0 Ω	CDN type Mx, Sx, Tx, AFx
Insertion loss (common mode)	8 ... 10.5 dB	150 kHz ... 230 MHz	0,4 dB	CDN adapter
AMN parameters:				
Impedance, Z_{CM} (common mode)	1 ... 65 Ω (Magnitude) -15° ... 90° (Phase)	9 kHz ... < 100 kHz 100 kHz ... 30 MHz 9 kHz ... 30 MHz	0,5 Ω 1,0 Ω 3,0 °	According CISPR 16-1-2 for V-AN (V-AMN, LISN)
Coupling factor (VDF) (common mode)	0 ... 50 dB	9 kHz ... < 30 kHz 30 kHz ... 30 MHz	0,6 dB 0,5 dB	
Decoupling factor (Isolation, common mode)	20 ... < 50 dB 50 ... 75 dB	9 kHz ... 30 MHz 9 kHz ... 30 MHz	4,0 dB 11,8 dB	
AAN parameters:				
Impedance, Z_{CM} (common mode)	5 ... 300 Ω (Magnitude) -40 ... 40° (Phase)	150 kHz ... 30 MHz 150 kHz ... 30 MHz	5,0 Ω 3,0°	According CISPR 16-1-2 for AAN (asymmetric artificial network, Y-network, unshielded symmetric signals) and AN (Artificial network for coaxial and other screened cables)
Coupling factor (VDF) (common mode)	0 ... 50 dB	150 kHz ... 30 MHz	0,5 dB	
Decoupling factor (Isolation, common mode)	20 ... < 50 dB 50 ... 75 dB	150 kHz ... 30 MHz 150 kHz ... 30 MHz	4,0 dB 11,8 dB	
EM clamp parameters:				
Impedance, Z_{CM} (common mode)	20 ... 1000 Ω	150 kHz ... < 80 MHz 80 kHz ... 230 MHz	20,0 Ω 40,0 Ω	According IEC 61000-4-6, Annex A (ABCD transformation) for EM and decoupling clamps
Coupling factor (VDF) (common mode)	0 ... 20 dB	150 kHz ... 230 MHz	1,7 dB	
Decoupling factor (Isolation, common mode)	0 ... 30 dB	150 kHz ... 230 MHz	3,0 dB	



SCS Directory

Accreditation number: SCS 0173

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
CMAD parameters:				
Insertion loss S21 (common mode)	0.05 ... 0.5	30 MHz ... 300 MHz	1,44 dB	According CISPR16-1-4, Section 8.5 and CISPR TR 16-3, Section 4.9.3.4 for Common Mode Absorbing Devices (CMAD)
Impedance S11 (common mode)	0.3 ... 1.0	30 MHz ... 300 MHz	1,58 dB	
SVSWR: TD-SVSWR acc. ANSI C63.25.1				According ANSI C63.25.1 for fully anechoic room (FAR) and semi anechoic chamber (SAC)
Time Domain (TD), frequency stepping 1 MHz	2,0 ... 10,0 dB 2,5 ... 10,0 dB	1 GHz ... < 6 GHz 6 GHz ... 18 GHz	1,46 dB 1,86 dB	Direct VNA measurement
Moving Average (MA), frequency stepping 120 MHz	1,5 ... 10,0 dB 2,0 ... 10,0 dB	1 GHz ... < 6 GHz 6 GHz ... 18 GHz	1,04 dB 1,47 dB	Moving Average (MA) and correlation according standard
Rod Antenna: Antenna factor (ECSM):	-100 ... 0 dB	9 kHz ... 36 MHz	0,42 dB	According CISPR 16-1-6, Clause 5.1
Current Probe parameters:				According CISPR 16-1-2, RTCA DO-160, MIL-STD-461, IEC 61000-4-6, ISO 11452-4
Insertion loss (common mode)	0,001 ... 10 Ω (Flat) 0,001 ... 10 Ω (Rise/Fall)	9 kHz ... 500 MHz 9 kHz ... 500 MHz	1,10 dB 1,59 dB	
Transfer impedance (common mode)	0,001 ... 10 Ω (Flat) 0,001 ... 10 Ω (Rise/Fall)	9 kHz ... 500 MHz 9 kHz ... 500 MHz	1,10 dB 1,59 dB	
AC Resistance: (Complex Impedance)	10 Ω ... 10 kΩ	3 Hz ... 150 kHz	2,5 %	According RF I-V measurement method
AC Capacitance: (Complex Impedance)	1 nF ... 100 µF	3 Hz ... 150 kHz	2,5 %	According RF I-V measurement method
AC Inductance: (Complex Impedance)	10 µH ... 1 H	3 Hz ... 150 kHz	2,5 %	According RF I-V measurement method

1) The dimensionless parts of the measurement uncertainty are relative values, related to the measured value.

* / * / * / * / *